



Pega Medical™

Hinge  
Plate™









For the treatment of pediatric  
angular deformities on long  
bones of arms and legs.

SURGICAL TECHNIQUE



The Hinge Pediatric Plating System™ is an articulated plate used to guide growth and deformity correction. Indicated in children and adolescents with deformities such as genu varum and genu valgum, it includes one plate and two screws. By fixing the plate on either side of the physal plate, axial growth can be controlled and angular deformity corrected.


Plates and screws are manufactured in medical grade Stainless Steel (316L as per ASTM F138). The plate is provided in three different sizes with distances between the anchoring points of the screws of 12, 16 and 20mm. The screws are provided in three different lengths of 25, 30 and 35mm.

Cat. #	Implant Components	
HPI012	Small Hinge plate 12 mm	
HPI016	Medium Hinge Plate 16 mm	
HPI020	Large Hinge Plate 20 mm	
HPS025	Screw 4.5 x 25 mm	
HPS030	Screw 4.5 x 30 mm	
HPS035	Screw 4.5 x 35 mm	

**Hinge Plate™**  
by Pega Medical

Developed in collaboration with:

Daniel Green, MD, FACS  
Pediatric Orthopedic and  
Scoliosis Services  
Hospital for Special Surgery, NY, NY USA




### Pre-Op Planning:


A careful preoperative planning and radiological analysis should be used to define the CORA, plane, direction and magnitude of the deformity intended to correct.

### Hinge Plate and Screws Selection:


The device is selected according to the patient's growth plate height and location to prevent invasion or damage of the physis. These variables define the plate size and screw length.




Initial location for placement of the Hinge Plate is achieved by inserting a guiding Guide Wire to localize the growth plate under fluoroscopic imaging (C-Arm). The Guide Wire should be inserted considering that the centerline of the Hinge Plate will be placed approximately at 6.5mm (1/4") from the Guide Wire




The Template is placed over the growth plate using the Croc Holder that is guided by the previously positioned Guide Wire.




Release the Template and verify proper seating over bone surface and alignment with growth plate through C-Arm




Define screw location by placing a second Guide Wire with the aid of the Dual Drill/Wire Guide at an angle away from the growth plate to avoid damage. Check position before drilling under C-Arm.



Using the cannulated drill bit, prepare screw hole to the depth indicated by the selected screw length.




Once the first hole is drilled, proceed with same method for second hole



Having both holes perforated proceed to remove the Template.




9



Proceed to position the Hinge Plate by using the Croc Holder assisted by the guiding Guide Wire.

10



Once the Hinge Plate is secured, the procedure is concluded by removing the guiding Guide Wire followed by the Croc Holder. Assure proper implant seating over the bone surface, avoiding periosteal damage, and check screw final position before closing.

## Surgical Instrumentation

Part name	Catalogue
Croc Holder	HPH100
Dual Guide	HPDG110
Template 12 mm	HPT112
Template 16 mm	HPT116
Template 20 mm	HPT120
Drill 3.0 mm	HPD130
Guide Wire	HPW150
Screw Driver	HPSD160
Case	HPC200





# Pega Medical™

1111 Autoroute Chomedey, Laval, Quebec CANADA H7W 5J8


Phone: 450-688-5144 • Fax: 450-688-1977

info@pegamedical.com

www.pegamedical.com

© 2020 Pega Medical, Inc.

Distributed by



Pega Medical

## Pediatric Orthopedics at its Best

